


CERTIFICATE OF ANALYSIS TECHNOFLUOR FXIII Activity




Attachment 06 DQFB00.02


Temp.: TFG327.07

FB51C00.02


REF 5800200-RUO
 LOT FB51C00.02
 2027-02-28

Content	volume	Reagent	LOT	
2x	3mL	HFP: Technofluor FXIII Buffer 3mL	FBC51C0	2027-02-28
2x	2mL	HFP: Technofluor FXIII Substrate 2mL	FBB51C0	2027-02-28
2x	2mL	HFP: Technofluor FXIII Reagent 2mL	FBA51C0	2027-02-28

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	Control N LOT 1P31C01	60.0 - 100.0 %	81.7 %	OK
	Control A LOT 3P42C00	28.5 - 47.5 %	38.1 %	OK

Prepared by Michael KAFKA 

27. MRZ. 2025
Date

Approved by Smilja DURANOVIC 

27. MRZ. 2025
Date