

CERTIFICATE OF ANALYSIS TECHNOFLUOR FXIII Activity



Temp.: TF0327.07

Attachment 06 DQFB00.02

FB52C00.02

REF	5800200-RUO
LOT	FB52C00.02
	2027-07-31

Content	volume	Reagent	LOT	
2x	3mL	HFP: Technofluor FXIII Buffer 3mL	FBC52C0	2027-07-31
2x	2mL	HFP: Technofluor FXIII Substrate 2mL	FBB53C0	2027-09-30
2x	2mL	HFP: Technofluor FXIII Reagent 2mL	FBA52C0	2027-08-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	Control N LOT 1P42C01	85.5 - 142.5 %	115.7 %	OK
	Control A LOT 3P51C00	27.8 - 46.3 %	38.1 %	OK

Prepared by Lukas GRUBELNIG 02. OKT. 2025
Date

Approved by Lucas HOLOVICZ 03. OKT. 2025
Date