

TECHNOFLUOR FXIII Activity



Attachment 07 DQFB00.01

Temp.: TP0327.06

FB41C00.02

REF	5800200-RUO
LOT	FB41C00.02
	2026-03-31

Content	volume	Reagent	LOT	
2x	3mL	HFP: Technofluor FXIII Buffer 3mL	FBC41C0	2026-03-31
2x	2mL	HFP: Technofluor FXIII Substrate 2mL	FBB41C0	2026-03-31
2x	2mL	HFP: Technofluor FXIII Reagent 2mL	FBA41C0	2026-03-31

Analytical Procedures		Requirement	Result	
visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Recovery of controls	Control N LOT 1P31C00	60.0 - 100.0 %	85.3 %	OK
	Control A LOT 3P23C01	25.5 - 42.5 %	31.7 %	OK

Prepared by Michael KAFKA

16. APR. 2024
Date

Approved by Suzana GRESAK PODAKAR

17. APR. 2024
Date