corgenix

CERTIFICATE OF PRODUCT CONFORMANCE

COAGULATION CONTROL

Coagulation Control 1

Product Code: 701-001
Lot Number: CC1 – 054*
Expiration Date: 2026-06-30

Reference Values
Coagulation Control 1

	Value (%)	Range (%)
Protein C Antigen	95%	76 - 114%
Protein S Antigen, Total	107%	86 - 129%
Protein S Antigen, Free by PEG Precipitation	90%	72 - 108%
Protein S Antigen, Free by Monoclonal Antibody	109%	88 - 131%
vWF Antigen	107%	85 - 128%
vWF Activity	89%	71 - 106%

* The alpha character suffix on the vial label lot number is for Corgenix internal use only; it is not shown on the value assignment sheet or this COC.

The information contained in this certificate has been provided by the manufacturer and is believed to be truthful and accurate. The lot listed above has been tested by Corgenix QC methods and meets all final specifications. This product is suitable for use until the stated expiration date. This product contains human source material; it has been tested at the donor level for HBsAg and antibodies to HIV1, HIV2, and HCV and has been found to be non-reactive.

Signed: _____ **Date:** 2024-02-26
Quality Assurance Specialist

Corgenix, Inc.
11575 Main Street, Suite 400
Broomfield, CO 80020
Telephone: (303) 457-4345
Telephone: (800) 729-5661
Fax: (303) 457-4519