



Certificate of Analysis

REAADS
VWF Activity ELISA

Kit Lot: VA-131
Plate Lot #: VAW-056
Product #: 10826
Expiration Date: 2027-06-02

Quality Control Results

VWF Activity		Testing Date: 2025-12-11		
Attribute	Value	Specification	Result by Assay	
Reference Plasma 100% OD	91 % of Normal	≥ 1.0	1.307	
Reagent Background O.D.		≤ 0.100	-0.001	
Reference Curve R-Squared		≥ 0.95	0.99358	
Control Recovery	Label Range	Result by Assay		
Normal Control	71 - 106 %	101.2 %		
Deficient Control	22 - 33 %	31.8 %		

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix, Inc.

Human sera used to prepare the controls included in this kit have been tested and shown to be nonreactive for HBsAG and negative for antibodies to HCV and HIV 1 and 2 by FDA required tests. However, all human blood derivatives, including patient samples, should be handled as though capable of transmitting disease.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

Quality Assurance Approval By: Souda Truong

Signed:  2025-12-15

Title: Quality Assurance Specialist

Corgenix, Inc.
11575 Main Street, Suite 400
Broomfield, CO 80020
Phone: (303) 457-4345
Fax: (303) 457-4519