

CERTIFICATE OF ANALYSIS

CHROMOGENIX

 **Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	Antithrombin 1 x 25 IU	EXPIRATION DATE: 2026-08-31
PRODUCT NUMBER:	81079639	
LOT NUMBER:	N0844864	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Antithrombin Activity	25	30	IU/vial	29	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0028	PASS
Intercept @ 37°C	0.40	0.60		0.47	PASS
r ²	0.990	--		0.999	PASS

PREPARED BY: Monika Skowronska Monika Skowronska
 PRINT SIGNATURE

DATE: 8/28/2024 8/28/24

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.M. Patel
 PRINT SIGNATURE

DATE: 8/28/2024 8-28-24

TITLE: Senior Group Leader