

CERTIFICATE OF ANALYSIS

CHROMOGENIX werfen

| | | | |
|-----------------|-------------------|------------------|------------|
| PRODUCT NAME: | FXa, 10 x 71 nkat | EXPIRATION DATE: | 2027-06-30 |
| PRODUCT NUMBER: | 00082098539 | | |
| LOT NUMBER: | N1248557 | | |

| TEST DESCRIPTION | SPECIFICATIONS | | UNITS | RESULTS | DISPOSITION |
|----------------------------------|----------------|-------|-----------|---------|-------------|
| | MIN | MAX | | | |
| FACTOR Xa, 71 nkat (650515) | | | | | PASS/FAIL |
| ACTIVITY | 61 | 81 | nkat/vial | 71 | PASS |
| STANDARD CURVE (r ²) | 0.990 | 1.000 | -- | 1.000 | PASS |
| CONTROL RECOVERY | 66.7 | 73.7 | nkat/vial | 73.1 | PASS |

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

PREPARED BY: Maria Somoza
 PRINT

Maria Somoza
 SIGNATURE

DATE: 01/21/25
 1/21/2025

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel
 PRINT

N.M. Patel
 SIGNATURE

DATE: 1-21-25
 1/21/2025

TITLE: Senior Group Leader