

CERTIFICATE OF ANALYSIS

CHROMOGENIX

 **Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	S-2302, 25 mg Substrate	EXPIRATION DATE:	2028-02-29
PRODUCT NUMBER:	82034039		
LOT NUMBER:	N1046682		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	23.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Maria Somoza
PRINT

Maria Somoza
SIGNATURE

DATE: 11/1/24

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel
PRINT

N.M. Patel
SIGNATURE

DATE: 11-1-24

TITLE: Senior Group Leader