

CERTIFICATE OF ANALYSIS

CHROMOGENIX werfen

PRODUCT NAME:	Antithrombin 10 x 10 IU	EXPIRATION DATE:	2027-05-31
PRODUCT NUMBER:	00082072039		
LOT NUMBER:	N1147571		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
ANTITHROMBIN, 10 IU (654023)					
ANTITHROMBIN ACTIVITY	9	11	IU/vial	11	PASS

All human source material used to manufacture this product have been tested according to FDA guidelines and found to be non-reactive for: HBsAg, antibodies to HIV1 and 2, and antibodies to HCV.

PREPARED BY: Brianna Randazzo *Brianna Randazzo* DATE: 12/9/2024
 PRINT SIGNATURE

TITLE: QC Analyst II
 APPROVAL SIGNATURE: Nehal Patel *N.m. Patel* DATE: 12-9-24
 PRINT SIGNATURE

TITLE: Senior Group Leader