

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	S-2288, 25 mg Substrate	EXPIRATION DATE:	2026-01
PRODUCT NUMBER:	82085239		
LOT NUMBER:	N0138567		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.2	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Monika Skowronska *Monika Skowronska* DATE: 1/26/2023
 PRINT SIGNATURE

TITLE: QC Analyst I

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza* DATE: 1/26/2023 *1/26/2023*
 PRINT SIGNATURE

TITLE: Senior Group Leader