

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company

PRODUCT NAME:	S-2366, 25 mg Substrate	EXPIRATION DATE:	2028-10
PRODUCT NUMBER:	82109039		
LOT NUMBER:	N1237485		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Fabiha Bushra
PRINT

Fabiha Bushra
SIGNATURE

DATE: 1/2/2024 1/2/24

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel
PRINT

N.M. Patel
SIGNATURE

DATE: 1/2/2024 1-2-24

TITLE: Senior QC Analyst