

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company

PRODUCT NAME:	S-2366, 25 mg Substrate	EXPIRATION DATE:	2029-06-30
PRODUCT NUMBER:	82109039		
LOT NUMBER:	N0845152		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Monika Skowronska
PRINT

Monika Skowronska
SIGNATURE

DATE: 9/19/2024 9-19-24

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel
PRINT

N.M. Patel

SIGNATURE

DATE: 9/19/2024 9-19-24

TITLE: Senior Group Leader