

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2765, 25 mg Substrate	EXPIRATION DATE:	2027-10-31
PRODUCT NUMBER:	82141339		
LOT:	N1046684		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.3	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.2	PASS

PREPARED BY: Fabiha Bushra *Fabiha Bushra* DATE: 11/1/2024 *11/1/24*
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N. m. Patel* DATE: 11/1/2024 *11-1-24*
 PRINT SIGNATURE

TITLE: Senior Group Leader