CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation Laboratory
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE:	2026-11-30
PRODUCT NUMBER :	82258563		
LOT NUMBER:	N1248565		

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	PASS/FAIL
		MIN	MAX	UNITS		
HIGH RANGE						
	Slope	0.33	0.57	--	0.49	PASS
	Intercept	---	0.12	--	0.10	PASS
	r	0.990	---	--	0.997	PASS
LOW RANGE						
	Slope	2.8	6.9	--	4.6	PASS
	Intercept	---	0.21	--	0.15	PASS
	r	0.990	---	--	0.997	PASS

KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N1248563	3 mL
COAMATIC FVIII S-2765 + I-2581	N1248564	6 mL
COAMATIC FVIII BUFFER	N1147987	24 mL

· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Monika Skowronska Monika Skowronska DATE: 1/22/2025 1-22-25
PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N. M. Patel DATE: 1/22/2025 1·22·25
PRINT SIGNATURE

TITLE: Senior Group Leader

QCM3082258563/49730503-1, Rev 03

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000