

CERTIFICATE OF ANALYSIS

CHROMOGENIX

 **Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE:	2027-04-30
PRODUCT NUMBER :	82258563		
LOT NUMBER:	N0452087		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.57	--	0.48	PASS
Intercept	---	0.12	--	0.10	PASS
r	0.990	---	--	0.997	PASS
LOW RANGE					
Slope	2.8	6.9	--	5.4	PASS
Intercept	---	0.21	--	0.15	PASS
r	0.990	---	--	0.995	PASS

KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0451795	3 mL
COAMATIC FVIII S-2765 + I-2581	N0452086	6 mL
COAMATIC FVIII BUFFER	N0451793	24 mL

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Monika Skowronska Monika Skowronska
PRINT SIGNATURE

DATE: 5/14/2025 5-14-25

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.M. Patel
PRINT SIGNATURE

DATE: 5/14/2025 5-14-25

TITLE: Senior Group Leader