

CERTIFICATE OF ANALYSIS

CHROMOGENIX



| | | | |
|------------------|----------------------|------------------|------------|
| PRODUCT NAME : | COAMATIC FACTOR VIII | EXPIRATION DATE: | 2027-04-30 |
| PRODUCT NUMBER : | 82258563 | | |
| LOT NUMBER: | N0452087 | | |

| TEST DESCRIPTION | SPECIFICATIONS | | UNITS | RESULTS | PASS/FAIL |
|------------------|----------------|------|-------|---------|-----------|
| | MIN | MAX | | | |
| HIGH RANGE | | | | | |
| Slope | 0.33 | 0.57 | -- | 0.48 | PASS |
| Intercept | --- | 0.12 | -- | 0.10 | PASS |
| r | 0.990 | --- | -- | 0.997 | PASS |
| LOW RANGE | | | | | |
| Slope | 2.8 | 6.9 | -- | 5.4 | PASS |
| Intercept | --- | 0.21 | -- | 0.15 | PASS |
| r | 0.990 | --- | -- | 0.995 | PASS |

KIT COMPONENTS

| NAME | LOT NUMBER | VOLUME |
|--------------------------------|------------|--------|
| COAMATIC FVIII FACTOR REAGENT | N0451795 | 3 mL |
| COAMATIC FVIII S-2765 + I-2581 | N0452086 | 6 mL |
| COAMATIC FVIII BUFFER | N0451793 | 24 mL |

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Monika Skowronska *Monika Skowronska*
 PRINT SIGNATURE

DATE: 5/14/2025 *5-14-25*

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel*
 PRINT SIGNATURE

DATE: 5/14/2025 *5-14-25*

TITLE: Senior Group Leader