

CERTIFICATE OF ANALYSIS



PRODUCT NAME :	<u>COAMATIC FACTOR VIII</u>	EXPIRATION DATE: <u>2027-12-31</u>
PRODUCT NUMBER :	<u>82258563</u>	
LOT NUMBER:	<u>N0169896</u>	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.57	--	0.53	PASS
Intercept	---	0.12	--	0.10	PASS
r	0.990	---	--	0.998	PASS
LOW RANGE					
Slope	2.8	6.9	--	6.5	PASS
Intercept	---	0.21	--	0.12	PASS
r	0.990	---	--	0.998	PASS

KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N1258735	3 mL
COAMATIC FVIII S-2765 + I-2581	N0169613	6 mL
COAMATIC FVIII BUFFER	N1258734	24 mL

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Monika Skowronska *Monika Skowronska* DATE: 2/5/2026 *2-5-26*
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel* DATE: 2/5/2026 *2-5-26*
 PRINT SIGNATURE

TITLE: Senior Group Leader