

# CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

Instrumentation  
Laboratory  
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE:	2027-04-30
PRODUCT NUMBER :	82258563		
LOT NUMBER:	N0552880		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.48	PASS
Intercept	---	0.12	--	0.08	PASS
r	0.990	---	--	1.000	PASS
LOW RANGE					
Slope	2.8	6.9	--	6.1	PASS
Intercept	---	0.21	--	0.11	PASS
r	0.990	---	--	0.999	PASS

## KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0552878	3 mL
COAMATIC FVIII S-2765 + I-2581	N0452456	6 mL
COAMATIC FVIII BUFFER	N0452458	24 mL

I have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Monika Skowronska Monika Skowronska DATE: 5-29-25  
PRINT SIGNATURE 5/29/2025

TITLE: QC Analyst II DATE: 5-29-25

APPROVAL SIGNATURE: Nehal Patel DATE: 5/29/2025  
PRINT SIGNATURE 5/29/2025

TITLE: Senior Group Leader