

CERTIFICATE OF ANALYSIS

CHROMOGENIX



PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE:	2028-05-31
PRODUCT NUMBER :	82258563		
LOT NUMBER:	N0564104		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.50	PASS
Intercept	---	0.12	--	0.11	PASS
r	0.990	---	--	0.994	PASS
LOW RANGE					
Slope	2.8	6.9	--	6.3	PASS
Intercept	---	0.21	--	0.14	PASS
r	0.990	---	--	0.999	PASS

KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0563752	3 mL
COAMATIC FVIII S-2765 + I-2581	N0563755	6 mL
COAMATIC FVIII BUFFER	N0563750	24 mL

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Monika Skowronska *Monika Skowronska*
 PRINT SIGNATURE

DATE: 6-8-26
6/8/2026

TITLE: QC Analyst III

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel*
 PRINT SIGNATURE

DATE: 6-8-26
6/8/2026

TITLE: Senior Group Leader