

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**

Instrumentation  
Laboratory  
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2027-10-31
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N1158414	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.47	PASS
Intercept	---	0.12	--	0.09	PASS
r	0.990	---	--	0.998	PASS
LOW RANGE					
Slope	2.8	6.9	--	5.1	PASS
Intercept	---	0.21	--	0.11	PASS
r	0.990	---	--	0.998	PASS

**KIT COMPONENTS**

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N1057798	3 mL
COAMATIC FVIII S-2765 + I-2581	N1158160	6 mL
COAMATIC FVIII BUFFER	N1157998	24 mL

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Monika Skowronska *Monika Skowronska* DATE: 12-9-25  
PRINT SIGNATURE

TITLE: QC Analyst II  
 APPROVAL SIGNATURE: Nehal Patel *N. M. Patel* DATE: 12-9-25  
PRINT SIGNATURE

TITLE: Senior Group Leader