CERTIFICATE OF ANALYSIS



PRODUCT NAME: PRODUCT NUMBER: COATEST APC Resistance Control Plasma Level 1

00082265063

LOT

N0341018

EXPIRATION DATE: 2026-03

Analysis with COATI	EST APC R V:					
APT Time	<u>CaCl₂</u>	1 12	39	seconds	33	PASS
	APC/CaCl ₂	75		seconds	90	PASS
APC Ratio		2.7	3.2		2.7	PASS

We hereby certify that the above product(s):

 have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;

- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY:	Monika Skowronska PRINT	Honika	SKOPRONSKO	_ DATE:	4/23/2024 4-23-24	
TITLE:	QC Analyst II					
APPROVAL SIGNATURE:		N.m. Patel		DATE:	4/23/2024 4 - 23 - 24	
	PRINT		SIGNATURE			
TITLE:	Senior QC Analyst					