

CERTIFICATE OF ANALYSIS



PRODUCT NAME:	CoaTest SP4 FVIII Kit	EXPIRATION DATE:	2026-07-31
PRODUCT NUMBER:	82409463		
LOT NUMBER:	N0744676		

TEST DESCRIPTION		SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
		MIN	MAX			PASS/FAIL
Coatest SP4 Factor VIII Kit (82409463)						
High Range	Slope	0.0077	0.0112	--	0.0087	PASS
	Intercept	--	0.250	--	0.137	PASS
	Standard Curve (r)	0.99	--	--	0.997	PASS
Low Range	Slope	0.020	0.080	--	0.043	PASS
	Intercept	--	0.406	--	0.232	PASS
	Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS:

DESCRIPTION	LOT NUMBER	VOLUME
COATEST SP FVIII CACL2	N0744127	6 mL
COATEST SP FVIII FACTOR IXa + FACTOR X	N0744322	3 mL
COATEST SP FVIII S-2765 + I-2581	N0744456	12 mL
COATEST SP FVIII BUFFER	N0744131	20 mL
COATEST SP FVIII PHOSPHOLIPID	N0744302	2 mL

PREPARED BY: Monika Skowronska Monika Skowronska
PRINT SIGNATURE

DATE: 8/26/2024 8-26-24

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.m. Patel
PRINT SIGNATURE

DATE: 8/26/2024 8-26-24

TITLE: Senior Group Leader

QCM0082409463-4, Rev. 02