



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ9600.01

Att.03

9652C00.02

Ceveron Cleaning Solution 25 mL

REF 9820300-RUO

LOT 9652C00.02

🕒 2026-10-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Cap color	white	confirmed	OK
	Vial type	white plastic vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Appearance	clear liquid	confirmed	OK
Kontrollwiederfindung Fibrinogen am Ceveron alpha	Control N LOT 1P53C00	2.53 - 3.79 g/L	3.40 g/L	OK
Recovery of controls Fibrinogen on Ceveron alpha	Control A LOT 3P54C00	0.85 - 1.27 g/L	1.13 g/L	OK

Durchgeführt von
Prepared by
Marcin MANUGIEWICZ

03. FEB. 2026

Datum / Date:

Freigabe von
Approved by
Maylis BAUSCHER

03. FEB. 2026

Datum / Date: