

CERTIFICATE OF ANALYSIS
Ceveron Wash Solution 30 mL



Attachment 03 DQ9200.03

9253C00.01

Temp.: TP0326.07

REF	9820320-RUO
LOT	9253C00.01
⌘	2030-06-30

Analytical Procedures		Requirement	Result	
visual inspection	Cap color	white	confirmed	OK
	Vial type	white plastic vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Appearance	clear liquid	confirmed	OK
pH		2.0 - 2.6	2.2	OK

Prepared by

DL
Lukas GRUBELNIG

27. OKT. 2025
 Date

Approved by

ikret NASUFI Fibret Nasuf

26. OKT. 2025
 Date