

CERTIFICATE OF ANALYSIS



PRODUCT NAME:	<u>COATEST APC Resistance Control Plasma Level 1</u>	EXPIRATION DATE: <u>2026-03</u>
PRODUCT NUMBER:	<u>00082265063</u>	
LOT	<u>N0341018</u>	

Analysis with COATEST APC R V:						
APT Time	<u>CaCl₂</u>	--	39	seconds	33	PASS
	<u>APC/CaCl₂</u>	75	--	seconds	90	PASS
APC Ratio		2.7	3.2	--	2.7	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Monika Skowronska Monika Skowronska DATE: 4/23/2024 4-23-24
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel N.m. Patel DATE: 4/23/2024 4-23-24
 PRINT SIGNATURE

TITLE: Senior QC Analyst