## **CERTIFICATE OF ANALYSIS**



100 Instrumentation Laboratory A Werfen Company

COATEST APC Resistance Control Plasma Level 2 00082286863 PRODUCT NAME: **EXPIRATION DATE: 2026-01** PRODUCT NUMBER: LOT NUMBER: N0148730

TEST DESCRIPTION Analysis with COATEST APC R V:		SPECIFIC	SPECIFICATIONS		RESULTS	DISPOSITION
		MIN	MAX	UNITS		PASS/FAIL
APT Time	CaCl <sub>2</sub>	28	41	seconds	36	PASS
APC Ratio		1.4	1.9		1.7	PASS

We hereby certify that the above product(s):

PREPARED BY:	PRINT	lonika Skowronska	Honka Sign	Skouronskol Ature	DATE:	2/16/2024 2-16-24	
TITLE:		QC Analyst II	••••••••••••••••••••••••••••••••••••••				
APPROVAL SIGNATURE:	E: Nehal Patel		to. m-Parel		DATE:	2-16-24	
	PRINT		\$IGN/	ATURE	_,,,_,	2102024	
TITLE:		Senior QC Analyst					

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.