



Certificate of Analysis

Factor Xla Calibrator

REF	Catalog number	1199
LOT	Batch code	CAW2635
	Use by	2029 08
	Upper limit of temperature	-20°C


Activity after reconstitution with 4.0 mL water:

83 mIU/mL

The above Factor Xla activity apply only to Lot#**CAW2635**.

The calibration has been made against the 1st International standard for Activated Blood Coagulation Factor XI (FXIa), Human, NIBSC code 13/100 using the Rox Factor Xla kit Ref 110050.

This Lot# has been approved for release after analyses according to established protocols with results demonstrating fulfillment of product specifications. All data concerning materials used, production data and release testing are documented and traceable under the individual Lot# which is stated on each package.



Pia Bryngelhed
Quality Manager, Rossix AB

October 1, 2024

Date



Rossix AB
Taljegardsgatan 3B
431 53 Molndal
Sweden

Phone: +46 31 706 8965
info@rossix.com
www.rossix.com