

Certificate of Analysis

REAADS

VWF Activity ELISA

Kit Lot: VA-126

Plate Lot #: VAW-052

Product #: 10826

Expiration Date: 2025-08-27

Quality Control Results

VWF Activity Testing Date: 2024-04-05			
Attribute	<u>Value</u>	Specification	Result by Assay
Reference Plasma 100% OD Reagent Background O.D.	82 % of Normal	≥ 1.0 ≤ 0.100	1.431 0.009
Reference Curve R-Squared		≥ 0.95	0.99399
Control Recovery	Label Range		Result by Assay
Normal Control	73 - 110 %		97.1 %
Deficient Control	17 - 28 %	*	26.3 %

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix,

Human sera used to prepare the controls included in this kit have been tested and shown to be nonreactive for HBsAG and negative for antibodies to HCV and HIV 1 and 2 by FDA required tests. However, all human blood derivatives, including patient samples, should be handled as though capable of transmitting disease.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

Quality Assurance Approval By: Souda Truong 2024-04-08 Title: Quality Assurance Specialist

Corgenix, Inc. 11575 Main Street, Suite 400 Broomfield, CO 80020 Phone: (303) 457-4345

Fax: (303) 457-4519

Controlled Document Page 1 of 1

COA048 00

Effective Date: 2012-08-01