## CERTIFICATE OF ANALYSIS





COATEST Heparin
25553963
N0441744

EXPIRATION DATE: 2026-08-31

TEST DESCRIPTION	SPECIFIC	SPECIFICATIONS		RESULTS	DISPOSITION	
	MIN	MAX	UNITS		PASS/FAIL	
A. FACTOR Xa						
ACTIVITY	- 61	81	nkat/vial	72	PASS	
STANDARD CURVE (r <sup>2</sup> )	0.990	1.000		1.000	PASS	
CONTROL RECOVERY	69.4	76.8	nkat/viat	75.1	PASS	
RESIDUAL MOISTURE (KF)	0	5	%	NA		
RESIDUAL MOISTURE (FMS)	0	12.0	Torr	0.3	PASS	

SUBSTRATE CONCENTRATION	14 25	16.2	mg/vial	15.15	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS
RESIDUAL MOISTURE	0	5	%	0.2	PASS

C. ANTITHROMBIN 10 IU, FOR COATEST HEPARIN					
ANTITHROMBIN ACTIVITY	9	11	IU/vial	10	PASS

D. NORMAL CONTROL PLASMA					
HEPARIN ANTAGONISTS	0	< 0.10	IU/mL	0.01	PASS
STANDARD CURVE:					
regression coefficient (r2)	0.99	1.00		0.99	PASS
slope	-0.800	-0.500		-0.666	PASS
intercept	0.650	1.030		0.825	PASS
CONTROL RECOVERY	0.4	0.6	IU/mL	0.6	· PASS
ABSORBANCE AT 405nm	0.00	0.400		0.360	PASS
APPEARANCE	yellowish-	l white, lyophiliz	ed powder	PASS	PASS

E. BUFFER pH 8.4 CONC., FOR COAT	EST HEPARIN				
APPEARANCE	clea	clear, colorless solution			PASS
pH @ 20 - 25°C	8.3	8.5		8.3	PASS

F. COATEST HEPARIN KIT					
SLOPE	-0.800	-0.500		-0.612	PASS
INTERCEPT	0.650	1.030		0.774	PASS
REGRESSION COEFFICIENT (r²)	0.99	1.00		0.99	PASS
CONTROL RECOVERY	0.4	0.6	IU/mL	0.6	PASS

KIT COMPONENTS:

NAME	LOT NUMBER	VOLUME
COATEST HEPARIN FACTOR Xa	N0249469	2 mL
COATEST HEPARIN S-2222	N0441733	4 mL
COATEST ANTITHROMBIN	N0249477	1 mL
COATEST HEPARIN NORMAL PLASMA	N0441731	1 mL
COATEST HEPARIN BUFFER	N0441736	12 mL

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY:	Monika Skowror	nska Honika Slupporska	DATE:	5/8/2024	5-8-24
	PRINT	SIGNATURE	_		
TITLE:	QC Analyst II				
APPROVAL SIGNATURE	Nehal Patel	p.m. Pure	DATE:	5/8/2024	5.8.24
	PRINT	SIGNATURE	-		
TITLE:	Senior QC Analy	st			