

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE:	2026-07-31
PRODUCT NUMBER :	82258563		
LOT NUMBER:	N0844863		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.49	PASS
Intercept	---	0.12	--	0.11	PASS
r	0.990	---	--	0.994	PASS
LOW RANGE					
Slope	2.8	6.9	--	5.0	PASS
Intercept	---	0.21	--	0.13	PASS
r	0.990	---	--	0.999	PASS

KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0844862	3 mL
COAMATIC FVIII S-2765 + I-2581	N0744319	6 mL
COAMATIC FVIII BUFFER	N0744018	24 mL

· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Monika Skowronska *Monika Skowronska* DATE: 8/29/2024 *8-29-24*
 PRINT SIGNATURE

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel *N.M. Patel* DATE: 8/29/2024 *8-29-24*
 PRINT SIGNATURE

TITLE: Senior Group Leader