

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	CoaTest SP4 FVIII Kit	EXPIRATION DATE:	2025-07
PRODUCT NUMBER:	82409463		
LOT NUMBER:	N0734040		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0099	PASS
Intercept	--	0.250	--	0.141	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.044	PASS
Intercept	--	0.406	--	0.222	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS:

DESCRIPTION	LOT NUMBER	VOLUME
COATEST SP FVIII CACL2	N0733812	6 mL
COATEST SP FVIII FACTOR IXa + FACTOR X	N0733814	3 mL
COATEST SP FVIII S-2765 + I-2581	N0734035	12 mL
COATEST SP FVIII BUFFER	N0734036	20 mL
COATEST SP FVIII PHOSPHOLIPID	N0734038	2 mL

PREPARED BY: Girle Reinoso
PRINT

G. Reinoso
SIGNATURE

DATE: 8/21/2023 8/21/23

TITLE: QC Analyst III

APPROVAL SIGNATURE: Nehal Patel
PRINT

N. M. Patel
SIGNATURE

DATE: 8/21/2023 8-21-23

TITLE: Senior QC Analyst