

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**



**Instrumentation  
Laboratory**

A. Werfen Company

PRODUCT NAME:	S-2238, 25 mg Substrate	EXPIRATION DATE:	2027-05-31
PRODUCT NUMBER:	82032439		
LOT NUMBER:	N0542548		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.4	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.2	PASS

PREPARED BY: Fabiha Bushra  
PRINT

Fabiha Bushra  
SIGNATURE

DATE: 5/30/2024 5/30/24

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel  
PRINT

N.M. Patel

SIGNATURE

DATE: 5/30/2024 5.30.24

TITLE: Senior Group Leader