

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

**Instrumentation  
Laboratory**  
A Werfen Company

PRODUCT NAME:	S-2251, 25 mg Substrate	EXPIRATION DATE:	2024-11
PRODUCT NUMBER:	82033239		
LOT:	N1117541		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	26.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Giulle Reinoso G. Reinoso  
 PRINT SIGNATURE

DATE: 12/8/2021 12/8/21

TITLE: QC Analyst III

APPROVAL SIGNATURE: Donna Tencza Donna Tencza  
 PRINT SIGNATURE

DATE: 12/8/2021 12/8/21

TITLE: Senior Group Leader