

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

**Instrumentation  
Laboratory**  
A Werfen Company

PRODUCT NAME:	S-2302, 25 mg Substrate	EXPIRATION DATE:	2027-08
PRODUCT NUMBER:	82034039		
LOT NUMBER:	N0441973		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Fabiha Bushra  
PRINT

*Fabiha Bushra*  
SIGNATURE

DATE: 5/14/24

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel  
PRINT

*N.M. Patel*  
SIGNATURE

DATE: 5.14.24

TITLE: Senior QC Analyst