

CERTIFICATE OF ANALYSIS



PRODUCT NAME:	S-2366, 25 mg Substrate	EXPIRATION DATE:	2024-12
PRODUCT NUMBER:	82109039		
LOT NUMBER:	N0202122		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.3	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.2	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 3-4-2020
 PRINT SIGNATURE

TITLE: QC Ananlyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 3-5-2020
 PRINT SIGNATURE

TITLE: QC Technical Supervisor