

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**  
Manufacturing Equipment

**Instrumentation  
 Laboratory**  
A Werfen Company

PRODUCT NAME:	S-2366, 25 mg Substrate	EXPIRATION DATE:	2028-10
PRODUCT NUMBER:	82109039		
LOT NUMBER:	N1237485		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Fabiha Bushra  
 PRINT

Fabiha Bushra  
 SIGNATURE

DATE: 1/2/2024 1/2/24

TITLE: QC Analyst II

APPROVAL SIGNATURE: Nehal Patel  
 PRINT

N.M. Patel  
 SIGNATURE

DATE: 1/2/2024 1-2-24

TITLE: Senior QC Analyst