

CERTIFICATE OF ANALYSIS

**CHROMOGENIX**



**Instrumentation  
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2403, 25 mg Substrate	EXPIRATION DATE:	2026-03
PRODUCT NUMBER:	82225439		
LOT NUMBER:	N0330584		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.1	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

PREPARED BY: Monika Skowronska Monika Skowronska DATE: 4/12/2023 4/12/23  
 PRINT SIGNATURE

TITLE: QC Analyst I

APPROVAL SIGNATURE: Dana Tenaga Dana Tenaga DATE: 4/12/23  
 PRINT SIGNATURE

TITLE: Senior Team Leader