




| | | |
|---|---|-----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ...00AZ |
|---|---|-----------|

TECHNOZYM® PAP Complex ELISA Kit 96T.

REF TC12060
LOT 120601C5.01
 2026-01-31

| Inhalt Content | Reagenz Reagent | LOT |  |
|-------------------|-------------------------|---|---|
| 1 x | PAP Plate | RGY41C0 | 2026-09-30 |
| 1 x 0.3mL | anti PAP POX | RBK41C0 | 2026-09-30 |
| 2 x 20mL | Wash buffer Concentrate | RBE41C0 | 2026-08-31 |
| 1 x 20mL | Dilution Buffer | RAF41C0 | 2026-09-30 |
| 1 x 12mL | TMB Substrate | RKN42C0 | 2026-08-31 |
| 1 x 12mL | Stop solution | RKK42C0 | 2026-09-30 |
| 1 x 0.5mL | PAP-Calibrator 1 | 1BM21C1 | 2026-01-31 |
| 1 x 0.5mL | PAP-Calibrator 2 | 2BM21C1 | 2026-01-31 |
| 1 x 0.5mL | PAP-Calibrator 3 | 3BM21C1 | 2026-01-31 |
| 1 x 0.5mL | PAP-Calibrator 4 | 4BM21C1 | 2026-01-31 |
| 1 x 0.5mL | PAP-Calibrator 5 | 5BM21C1 | 2026-01-31 |
| 1 x 0.5mL | PAP-high Control | ABM21C1 | 2026-01-31 |
| 1 x 0.5mL | PAP-low Control | BBM21C1 | 2026-01-31 |

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|------------------------|----------------------------|--------------------|----|
| OD | Calibrator 1 | OD: 1 – 2.5 | 1.40 | OK |
| | Calibrator 5 | OD: < 0.2 | 0.01 | OK |
| Verwendbarkeit Usability | High Control (ABM21C1) | 65.3 – 108.8 ng/mL | 89.35 | OK |
| | Low Control (BBM21C1) | 26.1 – 60.9 ng/mL | 40.47 | OK |


Durchführung von
Prepared by


Lucas HODONICZ

26. FEB. 2025

Datum / Date:

Freigabe von
Approved by


Stefan BRUCKBERGER

27. FEB. 2025

Datum / Date: