

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ...00AZ
---	---	-----------

TECHNOZYM® PAP Complex ELISA Kit 96T.

REF TC12060
LOT 120601C5.01
 2026-01-31

Inhalt Content	Reagenz Reagent	LOT	
1 x	PAP Plate	RGY41C0	2026-09-30
1 x 0.3mL	anti PAP POX	RBK41C0	2026-09-30
2 x 20mL	Wash buffer Concentrate	RBE41C0	2026-08-31
1 x 20mL	Dilution Buffer	RAF41C0	2026-09-30
1 x 12mL	TMB Substrate	RKN42C0	2026-08-31
1 x 12mL	Stop solution	RKK42C0	2026-09-30
1 x 0.5mL	PAP-Calibrator 1	1BM21C1	2026-01-31
1 x 0.5mL	PAP-Calibrator 2	2BM21C1	2026-01-31
1 x 0.5mL	PAP-Calibrator 3	3BM21C1	2026-01-31
1 x 0.5mL	PAP-Calibrator 4	4BM21C1	2026-01-31
1 x 0.5mL	PAP-Calibrator 5	5BM21C1	2026-01-31
1 x 0.5mL	PAP-high Control	ABM21C1	2026-01-31
1 x 0.5mL	PAP-low Control	BBM21C1	2026-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	OD: 1 – 2.5	1.40	OK
	Calibrator 5	OD: < 0.2	0.01	OK
Verwendbarkeit Usability	High Control (ABM21C1)	65.3 – 108.8 ng/mL	89.35	OK
	Low Control (BBM21C1)	26.1 – 60.9 ng/mL	40.47	OK

Durchführung von
Prepared by


Lucas HODOLICZ

26. FEB. 2025

Datum / Date:

Freigabe von
Approved by


Stefan BRUCKBERGER

27. FEB. 2025

Datum / Date: