



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQ...00AZ

**TECHNOZYM PAP Complex ELISA Kit 96T.**

**REF** TC12060

**LOT** 120601C6.01

**⌚** 2027-04-30

Inhalt Content	Reagenz Reagent	<b>LOT</b>	<b>⌚</b>
1 x	PAP Plate	RGY51C0	2027-08-31
1 x 0.3mL	anti PAP POX	RBK51C0	2027-08-31
2 x 20mL	Wash buffer Concentrate	RBE51C0	2027-10-31
1 x 20mL	Dilution Buffer	RAF51C0	2027-07-31
1 x 12mL	TMB Substrate	RKN53C0	2027-12-31
1 x 12mL	Stop solution	RKK61C0	2028-03-31
1 x 0.5mL	PAP-Calibrator 1	1BM51C0	2027-04-30
1 x 0.5mL	PAP-Calibrator 2	2BM51C0	2027-05-31
1 x 0.5mL	PAP-Calibrator 3	3BM51C0	2027-05-31
1 x 0.5mL	PAP-Calibrator 4	4BM51C0	2027-05-31
1 x 0.5mL	PAP-Calibrator 5	5BM51C0	2027-05-31
1 x 0.5mL	PAP-high Control	ABM51C0	2027-05-31
1 x 0.5mL	PAP-low Control	BBM51C0	2027-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	OD: > 0.7	0.9	OK
	Calibrator 5	OD: < 0.2	0.0	OK
Verwendbarkeit Usability	High Control (ABM51C0)	58.4 – 97.4 ng/mL	77.8	OK
	Low Control (BBM51C0)	23.0 – 53.6 ng/mL	32.9	OK

**Durchführung von  
Prepared by**

Lucas HOLOVICZ

22. APR. 2026

Datum / Date:

**Freigabe von  
Approved by**

Lukas GRUBELNIG

22. APR. 2026

Datum / Date: