



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS


DQ7BM0AZ

TECHNOZYM® PAP ELISA Control Set 2x0.5mL

REF TC12064

LOT 7BM41C0.02

 2026-01-31

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 0.5mL	PAP High Control	ABM21C1	2026-01-31
1 x 0.5mL	PAP Low Control	BBM21C1	2026-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit / Usability	High Control (ABM21C1)	65.3 - 108.8 ng/mL	43.83	OK
	Low Control (BBM21C1)	26.1 - 60.9 ng/mL	93.51	OK

Durchführung von  
Prepared by

Filoret NASUFI *Filoret Nasufi*

13. NOV. 2024

Datum / Date:

Freigabe von  
Approved by

*Lucas HOLMOCZ*

14. NOV. 2024

Datum / Date: