

TECO medical

TECO® Intact Proinsulin ELISA KIT

EN: Certificate of Analysis
ES: Certificado de análisis

DE: Analysezertifikat
I: Certificato di analisi

F: Certificat d'Analyse
PT: Certificado de análise

REF TE1011 RUO / TE1012 CE
LOT 0800025I



2027-02-12



EN: Refer to Product Instructions for Use for further Information

DE: Weitere Informationen finden Sie in der Packungsbeilage des Produkts

F: Se référer aux instructions d'utilisation du produit pour plus d'informations

ES: Consulte las Instrucciones de uso del producto para obtener más información

I: Fare riferimento alle Istruzioni per l'uso del prodotto per ulteriori informazioni

PT: Consulte as instruções de uso do produto para obter mais informações

CE in Europe
RUO in USA

EN: Standard Curve
Standards could vary from batch to batch

DE: Eichkurve
Standards können von Batch zu Batch unterschiedlich sein

F: Courbe Standard
La concentration des standards peut varier selon les lots

ES: Curva estándar
Los estándares pueden variar de un lote a otro

I: Curva standard
Gli standard possono variare da lotto a lotto

PT: Curva padrão
Os padrões podem variar de lote para lote

Standards	A pmol/L	$\bar{X}$ Extinction 450 nm	$\bar{X}$ Extinction 405 nm
A	0.0	0.000	0.000
B	2.6	0.101	0.031
C	8.2	0.332	0.176
D	19.6	0.802	0.412
E	45.0	1.763	0.996
F	130.4	-	1.466
L - Control 1	5.0 (3.3 - 6.8)	0.200	0.065
H - Control 2	21.2 (15.9 - 26.4)	0.863	0.276

ITEM

LOT

EXP

① Intact Proinsulin Antibody Coated Microtiter Plate	EBX-DZ-059	2027-02-28
② Blocking buffer	EBX-LA-051	2027-02-28
③ Antibody HRP Conjugate	EBX-AZ-059	2027-02-28
④ TMB Substrate	GZZ-IE-051	2028-01-31
⑤ Wash Solution	EBX-HZ-047	2027-02-28
⑥ Stop Solution	GZZ-KZ-034	2028-08-31
A Standard A	EBX-B0-040	2027-08-31
B Standard B	EBX-B1-040	2027-08-31
C Standard C	EBX-B2-040	2027-08-31
D Standard D	EBX-B3-040	2027-08-31
E Standard E	EBX-B4-040	2027-08-31
F Standard F	EBX-B5-040	2027-08-31
L Control 1	EBX-C1-040	2027-08-31
H Control 2	EBX-C2-040	2027-08-31

TECOmedical

Evaluation

Signature:

Name:

Function:

Date:

Daniel Lücking

Responsible Scientist

2025-09-18

QC Release

Signature:

Name:

Function:

Date:

Isabell Thomas

Quality Manager

2025-09-18