

**Technoview LMWH CAL - English****INTENDED USE**

Technoview LMWH (Low Molecular weight Heparin) CAL is a set of 5 calibration plasmas to be used for calibration LMWH measurements, optimized using Technochrom anti-Xa assay.

SUMMARY

Heparin is the most frequently used antithrombotic therapeutic drug. The biological activity of this sulfated glycosaminoglycan resides in its ability to accelerate (up to 2000-fold) the inhibitory effect of antithrombin (AT) on the coagulation proteases.

REAGENTS

The Technoview LMWH CAL contains:

	Reagent	Description
1 x 1 mL	Technoview LMWH CAL 1	Calibrator 1, human plasma, lyophilized, 0.0 IU/mL LMWH
1 x 1 mL	Technoview LMWH CAL 2	Calibrator 2, human plasma, lyophilized, ~0.4 IU/mL LMWH
1 x 1 mL	Technoview LMWH CAL 3	Calibrator 3, human plasma, lyophilized, ~0.8 IU/mL LMWH
1 x 1 mL	Technoview LMWH CAL 4	Calibrator 4, human plasma, lyophilized, ~1.2 IU/mL LMWH
1 x 1 mL	Technoview LMWH CAL 5	Calibrator 5, human plasma, lyophilized, ~1.6 IU/mL LMWH

Please consult the lot specific batch table provided with the controls for the exact concentrations in IU/mL of each calibrator.

Material required (not supplied with the kit)

- Distilled water
- Precision pipettes (1000 µL)
- Laboratory timer
- **REF** 5340250 Technochrom anti-Xa
- **REF** 5090042 Technoview LMWH CON L
- **REF** 5090044 Technoview LMWH CON M
- **REF** 5090046 Technoview LMWH CON H

Warning and precautions

- RUO for research use only.
- This kit is intended for use by personnel trained in laboratory procedures and universal precautions for the use of chemicals and potentially biohazardous substances must be applied.
- All human blood or plasma products as well as test samples must be considered as potentially infectious. They have to be handled with appropriate care and in strict observance of safety regulations. The rules pertaining to disposal are the same as applied to disposing hospital waste.
- Calibrators and control plasmas are made from human blood and any individual plasma involved in the procedure is tested HbsAg, HIV 1/2 Ab and HCV-Ab-negative. However, all human blood products should be handled as potentially infectious material.
- Get a Material Safety Data Sheet for this product from www.technoclone.com.

Stability and storage

The expiry date printed on the labels is only applicable to storage of the unopened containers at 2...8 °C.

Stability opened/ in use:

Reagent	18...25 °C (open vial)	2...8 °C (closed vial)	-20 °C
Technoview LMWH CAL 1	48 hours	7 days	1 month
Technoview LMWH CAL 2	48 hours	7 days	1 month
Technoview LMWH CAL 3	48 hours	7 days	1 month
Technoview LMWH CAL 4	48 hours	7 days	1 month
Technoview LMWH CAL 5	48 hours	7 days	1 month

Calibrator material can only be frozen once, in its original vial.

TEST PROCEDURE**Preparation of calibrators**

Before starting the test, all the required components are to be brought to room temperature.

When reconstituting plasmas, mixing reagents or buffers avoid foaming.

- Technoview LMWH CAL 1 - CAL 5: Reconstitute each calibrator vial with 1.0 mL of distilled water. Allow the reconstituted material to stand at room temperature for 30 minutes followed by swirling of the vial before use.

Performance of the test

The Technoview LMWH Calibrator Set is always used in combination with the Technoview LMWH CON L, Technoview LMWH CON M and Technoview LMWH CON H and the Technochrom anti-Xa kit.

The LMWH concentration of calibrator 1 - 5 may vary from one lot to another but is clearly indicated in the lot specific batch table provided in the calibration set.

Calibrator 1, calibrator 2, calibrator 3, calibrator 4 and calibrator 5 are used to establish a calibration curve. Results below the limit of quantification (instrument specific) should be declared as <LoQ. Samples found with a concentration above calibrator 5 should be re-tested in an appropriate dilution.

Technoclone provides instrument specific application sheets, which contain analyser / assay specific handling and performance information.

STANDARDISATION

The Technoview LMWH CAL is directly traceable to an international reference standard.

LITERATURE

Please contact Technoclone www.technoclone.com or your local distributor.

EDITORIAL NOTE


This document is available in several languages. The translations have been done using the master document in English. In the event of doubts or discrepancies, the wording in the master document in English shall take precedence.



Manufacturer



Lot



Storage temperature



Research Use Only



Expiry date



Catalogue number