

# CaCTI Reagent for T-TAS 01



## FOR RESEARCH USE ONLY – NOT FOR USE IN DIAGNOSTIC PROCEDURES

### SUMMARY:

The CaCTI Reagent for T-TAS 01 is used with the AR Chip for T-TAS 01 or the HD Chip for T-TAS 01 and the T-TAS 01 instrument.

For additional details about the use of the CaCTI Reagent for T-TAS 01, please consult the AR Chip for T-TAS 01 or HD Chip for T-TAS 01 package inserts and T-TAS 01 AR Chip measurement Reference guide or HD Chip measurement Reference guide.

### REAGENTS AND MATERIALS PROVIDED:

Item	Contents	Catalog Number
CaCTI Reagent for T-TAS 01	1 tube (0.4 mL)	19004

### MAIN COMPONENTS:

CTI (Corn-derived trypsin inhibitor) , CaCl<sub>2</sub>

### WARNINGS AND PRECAUTIONS:

- For research use only.
- Carefully follow the instructions and procedures described in this package insert.
- Do not use products beyond the expiration date printed on the label.
- If a white precipitate is observed when stored at 4°C, centrifuge before use.
- This reagent contains 0.05% sodium azide. Please handle with care.
- Proper handling and disposal methods should be followed for the reagent or any containers with reagent in accordance with local, state and federal regulations.









### STORAGE AND HANDLING REQUIREMENTS:

The CaCTI Reagent for T-TAS 01 is stable when stored at 2-8 °C until the expiration date on the package label.

### ASSISTANCE:

For assistance, please contact your local distributor.

## DEFINITION OF SYMBOLS:

Symbol	Definition
 YYYY-MM-DD	Use-by date
	Consult Instructions for use
	Batch code
	Catalog number
	Temperature limit
	Manufacturer
	Number of contents
	Do not use if package is damaged



FUJIMORI KOGYO CO.,LTD.  
1-1-1 Koishikawa, Bunkyo-ku,  
Tokyo 112-0002 Japan